



Investigation of Low and Medium Energy X-Ray Calibrations for Survey Meters

S. P. Jollota, N. P. Nelson, L. A. DeWerd

Department of Medical Physics, University of Wisconsin-Madison
Medical Radiation Research Center

29th Annual Council on Ionizing Radiation Measurements and Standards



INTRODUCTION

- In 2005, additional security controls and regulations were implemented by the United States Nuclear Regulatory Commission (NRC) in order to prevent the misuse of radioactive materials¹.
- Radiation Survey Meters (RSMs) are calibrated at Accredited Dosimetry Calibration Laboratories (ADCLs) using a ¹³⁷Cs source, with a monoenergetic photon energy of 662 keV.
- RSM calibrations are performed using a ¹³⁷Cs source and varying shielding and distance from the source²:
 - Irradiations are performed at each individual operating range of the RSM, specifically at the 50% point.
 - Adjustments are made to each individual operating range potentiometer if results are different than the 50% value.
 - 20% and 80% values of each operating range exposures are conducted to ensure that deviation from the given value does not exceed 5%³.
- Additionally, New York City passed a law in 2019 that requires RSM calibrations for electronic brachytherapy sources at the source energy, 50 kVp.
- The work investigates the feasibility of using currently available NIST traceable low and medium energy x-ray beam qualities as a surrogate for ¹³⁷Cs to calibrate RSMs.

Beam Quality	Effective Energy (keV)	Tube Current (mA)	Air Kerma, 1m (mGy/sec)	Exposure, 1m (R/hr)	Exposure, 2m, 0.1 mA (R/hr)
UW40-M	19.8	20	1.32	540	0.68
UW50-M	22.4	25	2.02	830	0.83
UW60-M	26.9	25	1.80	738	0.74
UW80-M	33.5	25	1.82	748	0.75
UW100-M	42.1	25	1.82	748	0.75
UW120-M	49.9	25	2.37	971	0.97
UW150-M	67.0	20	2.17	893	1.12
UW200-M	99.8	15	1.93	793	1.32
UW250-M	145	12	1.63	671	1.40

Table 1. NIST Traceable Beam Qualities and Associated UW-ADCL Measured Air-Kerma-Rates

Monitor Chamber Measurements

- Monitor chamber measurements were taken for each beam quality input voltage and corrected for tube output versus the linear relationship between tube current and air-kerma-rate.
- Exposures were done with each RSM at an input tube current of 0.1 and 0.5 mA to investigate the energy response from 19.8 to 145 keV.

RESULTS

- Monitor chamber measurements showed a 25% increase in tube output at 40 kV and 0.1 mA inputs and a 38% decrease at 250 kV and 0.1 mA (Figure 1)
- Correction factors for tube output were applied to calculated exposure rates in Table 1 and an energy response plot was tabulated for each RSM at 0.1 and 0.5 mA tube current inputs (Figure 2).




Figure 1. Correction factors associated with each beam quality to correct for actual tube output




Figure 2. Energy response plots for each RSM with corrected exposure rates

CONCLUSIONS

- Further Monte Carlo work should be done to simulate 320 kV beam qualities with additional filtration to see if the air-kerma-rates can be lowered to meet all RSM operating ranges.
- Additional physical measurements will be required to validate that any beam quality with a 320 kV x-ray tube could overcome the energy dependence of Geiger Mueller type RSMs.

ACKNOWLEDGEMENTS

Thank you to the University of Wisconsin - Accredited Dosimetry Calibration Laboratory (UW-ADCL) customers, whose continuing patronage supports ongoing research at the Medical Radiation Research Center (UWMRRC).

REFERENCES

- Murphy MK, Kamen J. Deciding between an X-Ray and ¹³⁷Cs Irradiator – It's not just about Energy Spectra. Radiation Research 2019;192(5):493-506.
- American National Standards Institute. American National Standard for Radiation Protection Instrumentation Test and Calibration, Portably Survey Meters, ANSI N323AB. 2013.
- National Council on Radiation Protection and Measurements. Calibration of Survey Instruments used in Radiation Protection for the Assessment of Ionizing Radiation Fields and Radioactive Surface Contamination, NCRP Report No. 112. 1991.

METHODS

X-Ray Beams

- COMET MXR-320/26 – kV range (0-320 kV), mA range (0.1-30 mA)

- 9 Moderately Filtered NIST Traceable Beam Qualities (Table 1)

RSMs

- 2 Ionization Type Survey Meters
 - Fluke 451B (Air-Communicating) – Operating Range (0.1 mR/hr – 50 R/hr)
 - Fluke 451P (Pressurized) – Operating Range (0.1 mR/hr – 5 R/hr)

- 1 Geiger Mueller Type Survey Meter
 - Ludlem 14C – Operating Range (0.1 mR/hr – 2 R/hr)

Exposure Rate Calculations

- Air-kerma-rates associated with each beam quality in Table 1 were measured at the UW-ADCL

- Exposure rates were scaled by the inverse square law and the linear relationship between input tube current and air-kerma-rate:

$$X = X_{ref} \frac{Current_{meas}}{Current_{ref}}$$