

Investigation of Potential Fluorescence Influence in Bremsstrahlung X-Ray Techniques

*RK Piper and LE Myers
Pacific Northwest National Laboratory*

Presented at the
*16th Annual Meeting
Council on Ionizing Radiation Measurements
and Standards
October 22-24, 2007*

Goals

Collect input from others

Suggest alternative to current NIST
Performance Test methods


History / Background

▶ X-ray Capability Expansion

- ANSI/HPS N13.11-2001
- Added roughly 41 bremsstrahlung techniques
- Standard filter recipes (N13.11)
 - NIST (L, M, H, S)
 - ISO (HK, WS, NS, LK)
 - Some NIST-ISO similarities



Comparison of Energy Distribution for ~60 kVp ISO Techniques*




* Spectra obtained from Ankerhold, U. Catalogue of X-ray spectra and their characteristic data - ISO and DIN radiation qualities, diagnostic and therapy radiation qualities, unfiltered X-ray spectra -. PTB Bericht, PTB-Dos-34, April 2000, ISBN 3-89701-513-7.

Low Air-Kerma Rate (LK) Series

- ▶ LK series unique to ISO
- ▶ 30, 35, 55, 70, 100, 125, 170, 210, 240 keV
- ▶ Intensity (relative to beam current) is significantly lower than comparable energy H, NS techniques

Comparison of Beam Intensity for ~60 kVp ISO Techniques*





* Spectra obtained from Ankerhold, U. Catalogue of X-ray spectra and their characteristic data - ISO and DIN radiation qualities, diagnostic and therapy radiation qualities, unfiltered X-ray spectra -. PTB Bericht, PTB-Dos-34, April 2000, ISBN 3-89701-513-7.

LK Development


- ▶ New Techniques Characterized
 - Half-Value Layer
 - Homogeneity Coefficient
 - Uniformity
 - Spectra (rough qualitative view)
 - Calibrated w/NIST traceable Ion Chamber
- ▶ Implemented in 2002

Calibration/Performance Test Tools



- ▶ Calibration based on reference-class ion chambers
- ▶ Performance testing based on similar chambers

PM-30 Energy Response
NIST Calibration Coefficient - Narrow Spectra




LK Implementation

- ▶ Processors - evaluate algorithms
- ▶ Discontinuity with LK techniques
 - Response inconsistent with average energy
 - Calcium – based phosphors
 - 2 independent reports w/data


LK Implementation - Issues -

► Abnormal response indicated on Ca-based phosphor



► Courtesy of Stanford Dosimetry, LLC

► Courtesy of Mike Lantz, et.al.



LK Implementation - Issues -


- ▶ Discontinued use
 - 170, 210, 240
- ▶ Initiated Investigation
 - Earlier HVL challenges
 - HPGe Spectra
 - TLD Evaluation

HVL Measurement

- LK-240 -


- ▶ Reconsidered initial difficulty with HVL assessment
 - Initial HVL measurement ~25% high
 - Suggests higher energy
 - Initial HC measurement ~15% low
 - Suggests spectrum too broad

HVL Measurement (light filtration)



Filters for ~240-250 keV


HK-250



WS-250




NS-250



LK-240



HVL Measurement (Narrow Spectra)




► Reassessment w/shadow shield – better than 5%/10%

Spectrum Evaluation


- ▶ Low Efficiency HPGe
- ▶ Requires low beam current (~ 0.1 mA)
- ▶ For most – requires “pin-hole” collimator
 - Adds scatter and fluorescence
- ▶ Artifacts possibly induced by detector
 - Ge, In and normal scatter influences
- ▶ Calibration
 - Energy - Yes
 - Efficiency - No

LK-240 Bremsstrahlung Spectrum




* Spectra obtained from Ankerhold, U. Catalogue of X-ray spectra and their characteristic data - ISO and DIN radiation qualities, diagnostic and therapy radiation qualities, unfiltered X-ray spectra -. PTB Bericht, PTB-Dos-34, April 2000, ISBN 3-89701-513-7.

LK-240 Total Spectrum



* Spectra obtained from Ankerhold, U. Catalogue of X-ray spectra and their characteristic data - ISO and DIN radiation qualities, diagnostic and therapy radiation qualities, unfiltered X-ray spectra -. PTB Bericht, PTB-Dos-34, April 2000, ISBN 3-89701-513-7.


PNNL LK-240 - Low Energy Region




NIST TLD Intercomparison

- ▶ Used modified 8807 dosimeter
 - 2 x CaF_2
 - 2 x LiF
 - All filtered the same (plastic)
- ▶ Exposures conducted in air
- ▶ Wide beam aperture (NIST)
- ▶ Same air-kerma
- ▶ Same irradiation date
- ▶ Evaluated LK240 and NS250
 - Total signal on LiF
 - CaF_2 :LiF ratio


Intercomparison of NIST/PNNL X-rays Li-based Element Response




Intercomparison of NIST/PNNL X-rays Ca:Li Phosphor Ratio



X-ray System Components (Top View)




X-ray System Components (Top View)




X-ray Shield

- ▶ Steel construction
- ▶ Large aperture – keep penumbra from reflecting
- ▶ Reduce aperture as needed
- ▶ Use other materials (W, Pb, SS)



Summary of LK Collimator Studies


NIST, PNNL(7/2003), PNNL(2/2004)



X-ray Shield II


(May 2004)

- ▶ Earlier success limited
- ▶ Shield creating scatter?
- ▶ Limited to Pb sheets




Summary of LK Collimator Studies

NIST, PNNL(7/2003), PNNL(2/2004), PNNL (5/2004)



Initial Study PNNL Initial Study NIST W/ Pb Col. 7-03 W/Steel Col. 2-04 Pb Top-Bot 5-04

HVL Measurement (Narrow Spectra)



Summary

- ▶ Urgency curtailed for LK Techniques
 - Undesirable for passive dosimetry exposures
 - Anticipated removal from N13.11
 - Investigation invasive upon facility
- ▶ New applications may benefit
 - Low rate photon detector evaluations

Summary (Cont'd)

- ▶ Improve spectrum measurement capability
 - Use better collimators
 - Proficiency in measurement and striping capability
 - Efficiency calibration of spectrometer

- ▶ Once origin of “contamination” identified...
 - Isolate (via shielding)
 - If necessary recalibrate and recharacterize affected techniques
 - If system modifications are necessary – recalibrate entire inventory.

Summary (Cont'd)

- ▶ Develop intercomparison system
 - Energy sensitive
 - Adequate sensitivity to resolve subtleties
 - Active detector – immediate feedback
 - Interlaboratory evaluations
 - NIST/NVLAP performance testing
 - National Need?

Contacts

Kim Piper

(509)376-6187

kim.piper@pnl.gov

Lyn Myers

(509)376-8310

lyn.myers@pnl.gov