

# NIST X-RAY (10 kV to 300 kV) Air-Kerma Calibration Service

by  
Michelle O'Brien  
Physicist, Dosimetry Group

*The HOME of the US realization of the  
SI unit of Radiation Dose for x-rays*




*MQSA medical traceability*



*diagnostic and therapeutic radiology*



*dose traceability for radiation  
protection and personnel dosimetry*



**NIST Mission:** To promote U.S. innovation and industrial competitiveness by advancing measurement science, standards, and technology in ways that enhance economic security and improve our quality of life.

## Disclaimer

Certain commercial equipment, manufacturers, instruments, or materials are identified in this presentation in order to specify the experimental procedure adequately. Such identification is for informational purposes only and is not intended to imply recommendation or endorsement by the National Institute of Standards and Technology, nor is it intended to imply that the manufacturer, materials, or equipment are necessarily the best available for the purpose.

## 17025 Quality Management System

### **NIST QMI 4.1 Impartiality** 4.1.1

Impartiality NIST conducts its measurement services in a fair and open-minded manner that is free of conflict of interests, bias or prejudice ensuring impartiality in the operations and measurements of the laboratory.

**NIST QMI 4.2 Confidentiality** 4.2.1 NIST is legally obligated to provide protection of confidential or proprietary information obtained or created during the performance of laboratory measurement services activities.

Department of Commerce

NIST PML

Radiation Physics Division

# *Dosimetry Group*

**Mission:** To develop standards by realizing the SI unit of dose, the gray, and advance the measurement of quantities important in the radiological sciences through programs in the dosimetry of x rays, gamma rays, electrons and other charged particles.

$$1 \text{ Gy} \equiv 1 \text{ J / kg}$$




Radiation Physics Building





**NIST**  
**Gaithersburg, MD**  
**H-wing of the**  
**Radiation Physics**  
**Building 245**




**Radiation Physics Building**

# Calibration and Measurement Capabilities (CMCs)

The goal for **Medical Physics Applications** is to provide traceability of radiation dose to clinical end users to ensure safety and efficacy of diagnostic and therapeutic procedures that involve ionizing radiation.

- 17025 and 17043 Quality management system
- BIPM.RI(I)-K2: Low energy x ray (10 kV to 50 kV)
- BIPM.RI(I)-K3: Medium energy x ray (50 kV to 300 kV)
- BIPM.RI(I)-K7: Mammography
- AAPM ADCL traceability and TG-61
- ISO 4037: Reference Radiation x-ray beams
- FDA-MQSA
- ANSI: personnel dosimetry performance standards, DHS






**Control Room**




Equipment Room




X-Ray Range




100 kV



300 kV



mammography



NIST X-Ray Range

Three sub-basement spaces became one first floor facility



W 300 kV



Identical shutter W disk  
150 msec


W 100 kV




Mo & Rh



# Air-Kerma Rate (Gy/s) Measurement



# Dosimetry of X Rays ( $E < 300$ keV)



$$K_{\text{air}} = Q_{\text{air}} \left( \frac{\overline{W}_{\text{air}}}{e} \right) \frac{1}{\rho_{\text{air}} V}$$

KERMA = Kinetic Energy Relaxed per unit MAss

Secondary electrons  liberated charge in a given mass of air

$$\text{C} \cdot \left[ 33.97 \frac{\text{J}}{\text{C}} \right] \cdot \frac{1}{\text{kg}} = \frac{\text{J}}{\text{kg}} = \text{Gy}$$


Air kerma can be measured absolutely with a free-air ionization chamber

# NIST Primary Free-Air Ionization Chambers

## Electronic Brachytherapy

Lamperti 10 kV to 60 kV




Air-path length 39 mm




Wyckoff-Attix 60 kV to 300 kV  
308 mm

\* Ritz 10 kV to 100 kV  
127 mm

\* Attix 10 kV to 50 kV  
212 mm



# The SI unit of dose, Gy, Air-Kerma Rate (Gy/s) as Realized by Free-Air Chambers



$$\dot{K} = \frac{I}{\rho_{\text{air}} V} \frac{W_{\text{air}}}{e} \frac{1}{1 - g_{\text{air}}} \prod_i k_i$$

$\dot{K}$  is the air-kerma rate at a given distance in air.


$I / (\rho_{\text{air}} V)$  is the measured ionization current divided by the mass of air in the measuring volume.




$W_{\text{air}}$  is the mean energy expended by an electron of charge  $e$  to produce an ion pair in dry air. The value used at NIST is  $W_{\text{air}}/e = 33.97 \text{ J/C}$ .


$g_{\text{air}}$  is the fraction of the initial kinetic energy of secondary electrons dissipated in air through radiative processes, which is 0.0 (negligible) for x rays with energies less than 300 keV.

$\prod k_i$  is the product of various correction factors.

Range of corrections: 0.983 to 1.13  
Air attenuation: 1.005 to 1.05











Two pair of test chamber positions

# Operating Region of Ionization Chambers





# NIST MQSA Beam Parameters



**CIRMS History: 1994 MPD A.1**

Beam code	Tube voltage (kV)	Filtration (mm)	Half-value layer (mm Al)
<b>Mo Anode</b>			
Mo/Mo23	23	0.032 Mo	0.288
Mo/Mo25	25	0.032 Mo	0.313
Mo/Mo28	28	0.032 Mo	0.346
Mo/Mo30	30	0.032 Mo	0.370
Mo/Mo35	35	0.032 Mo	0.404
Mo/Rh28	28	0.029 Rh	0.420
Mo/Rh32	32	0.029 Rh	0.453
Mo/Mo25x	25	0.030 Mo + 2.0 Al	0.551
Mo/Mo28x	28	0.030 Mo + 2.0 Al	0.589
Mo/Mo30x	30	0.030 Mo + 2.0 Al	0.633
Mo/Mo35x	35	0.030 Mo + 2.0 Al	0.715
<b>Rh Anode</b>			
Rh/Rh25	25	0.029 Rh	0.351
Rh/Rh30	30	0.029 Rh	0.438
Rh/Rh35	35	0.029 Rh	0.512
Rh/Rh40	40	0.029 Rh	0.559
Rh/Rh30x	30	0.029 Rh + 2.0 Al	0.814
Rh/Rh35x	35	0.029 Rh + 2.0 Al	0.898

# NIST H Series Beam Parameters

Beam code	Additional filtration				Half-value layer (HVL)	
	Al (mm)	Cu (mm)	Sn (mm)	Pb (mm)	Al (mm)	Cu (mm)
H10	0.105				0.051	
H15	0.5				0.16	
H20	1.01				0.36	
H30	4.50				1.2	
H40	4.53	0.26			2.93	
H50	4.0			0.1	4.16	0.14
H60	4.0	0.61			6.06	0.25
H100	4.0	5.2			13.51	1.15
H150	4.0	4.0	1.51		16.93	2.43
H200	4.0	0.6	4.16	0.77	19.72	4.1
H250	4.0	0.6	1.04	2.72	21.59	5.19
H300	4.1		3.0	5.0	23.55	6.19

**CIRMS History:**  
 1998 MPD A.5  
 M80 and M120  
 diagnostic x-ray

Beam Series  
 NIST L, M, & H  
 ISO 4037  
 CCRI – BIPM

Mammography: ( Mo/Mo, Mo/Rh, Rh/Rh)  
 Mammography with W/ ( Ag, Mo, Rh, Al)

# REPORT OF AIR-KERMA CALIBRATION

OF

Customer Name

Customer Address

Chamber Model and SN



## Calibration Conditions

**Chamber orientation:** The cavity was positioned in the center of the beam with the stem of the chamber perpendicular to the beam direction.

**Chamber potential:** 500 volts was applied to the chamber and negative charge was collected.

**Chamber rotation:** The high voltage wire faced the source of radiation.

**Background current:** The background current is less than 0.01 % of the exposure current.

**Current ratio:** The current ratio at the full to half collection potential is 1.003 using the S60 beam.

**Calibration date:** Date range

**Temperature range:** 294.2 K (21.1 °C) to 294.6 K (21.5 °C)    **Pressure range:** 99.8 kPa to 100.1 kPa

## RESULTS

Beam Code	Half-Value Layer		Calibration Coefficient (Gy/C) at 295.15 K (22 °C) and 101.325 kPa (1 Atm)	Air-Kerma Rate (Gy/s)	Beam Diameter (cm)	Calibration Distance (cm)
	mm Al	mm Cu				
M150	10.30	0.66	1.034E+06	4.48E-04	10	100
M200	14.73	1.64	1.042E+06	4.98E-04	10	100
M300	21.77	5.3	1.058E+06	1.84E-04	10	100
S60	2.81	0.09	1.037E+06	3.82E-04	10	100

## Expanded, Combined Uncertainty, in Percent


Beam Code	Air-Kerma Rate (Gy/s)	Calibration Coefficient (Gy/C)
M150	0.91	0.93
M200	0.92	0.94
M300	0.92	0.95
S60	0.92	0.94

The results relate only to the instrument calibrated in this report.

## X-ray air kerma uncertainties

Components	Type A	Type B
	<b>Primary standard</b>	
air density	0.01	0.07
FAC stdev charge	<b>0.02</b>	0.06
humidity		0.03
volume	0.04	0.01
g		0.02
<b>W/e (ICRU90)</b>		0.35
<b>ICRU90 kiikw</b>		0.05
air attenuation, $k_a$	0.05	0.02
electric field distortion		0.20
electron loss, $k_e$		0.10
penetration of aperture		0.04
penetration of chamber face		0.01
polarity difference	0.05	
recombination loss, $k_r$	0.10	
fluorescence $k_f$		0.03
scattered photons, $k_p$		0.07
<b>quadratic sum</b>	0.13	0.44
<b>combined standard uncertainty</b>	0.46	
<b>FAC expanded uncertainty</b>	<b>0.92</b>	
<b>Test Chamber</b>		
air kerma rate	0.13	0.44
air density	0.01	0.07
Test stdev charge	<b>0.06</b>	0.06
distance	0.01	
humidity		0.03
radiation background		
<b>quadratic sum</b>	0.14	0.45
<b>combined standard uncertainty</b>	0.47	
<b>Test expanded uncertainty</b>	<b>0.94</b>	

Calibration:  
chamber ships  
to NIST



NIST delivers  
calibration  
report with a  
Calibration  
Coefficient  
(Gy/C) and  
uncertainty to  
customer.

**REPORT OF PROFICIENCY TEST  
FOR AIR-KERMA CALIBRATIONS**

OF  
Customer Name  
Customer Address

Radiation Detection Chamber: NIST chamber Model and SN

Participant's Calibration Conditions as reported by Participant

Calibration distance:  
Chamber orientation:  
Chamber collection potential:  
Chamber rotation:  
Beam size:  
Nominal Air-Kerma rates:  
Uncertainty:

*Customer provided data*

Comparative Results for the NIST Transfer Standard

NIST Beam Code	NIST Calibration Coefficient 295.15 K (22 °C) and 101.325 kPa (1 Atm) (Gy/C)	Participant Calibration Coefficient 295.15 K (22 °C) and 101.325 kPa (1 Atm) (Gy/C)	Difference in Percent (%)
Mo/Mo23	NIST CC	Customer CC	-0.05
Mo/Mo25	NIST CC	Customer CC	-1.44
Mo/Mo28	NIST CC	Customer CC	-0.60
Mo/Mo30	NIST CC	Customer CC	-1.16
Mo/Mo35	NIST CC	Customer CC	0.33




**FDA MQSA**  
Calibration: +/- 6 %  
PT: +/- 3 %  
2-year frequency

Dosimetry Group Number DG:12345-YR  
NIST ID 46050S Order Number: 682.02/O-0000012345-YR  
Report Date: MM/DY/YR  
Page 4 of 13



# NIST technical support for the Mammography Quality Standards Act (MQSA)

Quantity	MQSA Final Regulation Citation	Regulation or Action Levels	NIST Support	
Beam quality (HVL)	900.12(e)(5)(iv)	Operating Voltage (kV)	Minimum HVL (mm Al)	17 reference beam qualities between 23 kVp and 40 kVp
		20	0.20	
		25	0.25	
		30	0.30	
Dose	900.12(e)(5)(vi)	<b>Cannot exceed 3.0 mGy (0.3 rad) per exposure</b>		Air-kerma calibration
Instrument calibration	900.12(e)(12)	Instruments used by medical physicists in their annual survey to measure the air kerma or air kerma rate from a mammography unit shall be calibrated at least once every 2 years and each time the instrument is repaired. The <b>instrument calibration</b> must be traceable to a national standard and calibrated with an accuracy of ±6 percent (95 percent confidence level) in the mammography energy range.		Air-kerma calibration and technical support to calibration facilities
Radiation output	900.12(e)(5)(x)	<b>Minimum output of 4.5 mGy/s</b> using Mo/Mo28 at any SID where the system is designed to operate with the detector center located 4.5 cm above the breast support surface with compression paddle in between. After October 28, 2002, the system, under the same measuring conditions shall be capable of producing a minimum output of 7.0 mGy/s.		Air-kerma calibration
Traceability	900.2(xx)	Instrument must be calibrated at either NIST or at a calibration laboratory that participates in a <b>proficiency test</b> with NIST at least once every 2 years and the agreement must be within <b>+/- 3 %</b> of the NIST value.		NIST proficiency test
kVp	900.12(e)(5)(ii)	<b>Cannot exceed +/- 5% of indicated kVp</b>		Proposed calibration service



SI  
BIPM

Uncertainty gain

NIST 0.9 %

Calibration Facility  
1.5 % to 4 %


Clinical

Traceability

A calibration provides traceability.

A proficiency test demonstrates ability to transfer air kerma but requires a linking calibration.

# Measurement Traceability for X-Ray Air-Kerma



# X-ray Air-Kerma Calibration Capabilities over the decades

